

BIDHAN CHANDRA KRISHI VISWAVIDYALAYA
COLLEGE OF AGRICULTURE
BurdwanSadar

Dr. Poly Saha
 Assistant Professor (Plant Pathology)



Agricultural Farm, Gate No-1,
 Kalna Road, Burdwan,
 Pin 713101, West Bengal
 Email: poly.saha@gmail.com
 Mobile: +919434586429

No. 54 COA/ BWN / BCKV 2025-26

Date:

NOTICE INVITING TENDER

The Principal Investigator of project titled "Evaluation of bio-efficacy of ALF-405 to control Sheath blight, Blast, Brown spot & Grain discoloration in Rice" implemented at College of Agriculture (Extended campus of BCKV), Agriculture Farm, Gate no 1, Kalna Road, PurbaBurdwan - 713101, West Bengal" under ad-hoc project (Code no. SCH-30873) sponsored by Atul Ltd is inviting sealed quotations from the competent and bonafidesuppliers/distributors/dealers/agents/vendors/manufacturers for supplying the equipment as per the specifications stated below with in **20 days from the date of publication in the website** upto 5 pm. Please mention terms and condition clearly against the item, if any.

Sl No.	Item	Specification	Quantity
1.	Laminar Air Flow	Size: 3ft* 2ft* 2 Ft Laminar cabinet inner chamber made of stainless-steel outer chamber made of MS with powder coating. High efficiency particular airfilter to achieve the air purification up to 0.3 microns in working area. Working area 3 ft * 2ft.* 2ft. Blower fitted with 1/4 HP motor with 1200 to 1350rpm. Pre filter made of high-grade nylon net fixed in SS frame for first stage air purification through blower system, closed inner chamber fitted with HEPA having very accurate performance rate of air filtration rated 99.99%, resulting in ceasing all air bour molecule of particle up to 0.3 micron in working area of laminar bench. Working area of laminar airflow cabinet illuminated by fluorescent light. Cabinet operated at 230volt single phase 50 Hz.AC supply. Fitted with UV germicidal lamp for sterilization. Fitted with acrylic front door sliding type. Fitted with manometer for measurement of HP filter choking system. Fitted with cock gas connection. Warranty: 3 years <i>(For bon motor & circuit)</i>	1

- i) Price: The price of the equipment should be quoted in net per unit (including taxes and duties etc.). Quoted rate must be for destination including packing, installation and delivery charges upto the laboratory at College of Agriculture, Burdwan, Extended Campus of BCKV, Kalna Road, PurbaBurdwan, WB-713101 with satisfactory installation and demonstration.
- ii) Supporting documents:
 - a) Bid papers should accompany authorization certificate from original manufacturer, trade license, GST registration, proprietary certificate (if any) etc.
 - b) Photocopy (Self attested) of the original supporting document in favour of the specification- claim for the item must have to be submitted.
 - c) User list along with certificate from reputed users also need to be submitted.
 - d) Photocopy of supporting document of assured after sale service in Eastern India and availability of spare parts need to be submitted.

Price bid of the vendors will be compared only if technical specificity as appended against the item is fulfilled. The Viswavidyalaya reserve the rights to accept or reject any tender without showing reason. The quotation reference number and subject must be mentioned on the top of envelope.

N.B: Please read the terms and condition carefully.

As per decision of the 82nd meeting of LTC held on 10/10/2025 tender notice is allowed to publish on per site.

Subrata Kumar
 10/10/25

Saha
 Principal Investigator

Dr. Poly Saha (Principal Investigator)
 Evaluation..... of ALF-405.....in Rice.
 Project Code: 30873
 Department of Plant Pathology
 College of Agriculture, Burdwan
 BCKV, Burdwan Sadar, WB 713101